

TEST REPORT



한국산업기술시험원
Korea Testing Laboratory

Report No. : 20-056955-02-1

Page of Pages : (1) / (2)



1. Client

Name : Korea Ozone · UV Association

Address : 130 dong, Seoul National University, 1, Gwanak-ro, Gwanak-gu, Seoul 08826, Korea

Date of Receipt : 2020. 09. 04.

2. Use of Report : For PA certification

3. Test Sample

Description : PLASMA AIR DISINFECTOR

Manufacturer : Health Well Medical Inc.

Model Name : HPA-36C

Serial Number : A036I08A001

Remark : ***

4. Date of Test : 2020. 11. 04. ~ 2020. 11. 05.

5. Location of Test :

☒ KTL Permanent Test Lab (Address : 87, Digital-ro 26-gil, Guro-gu, Seoul, KOREA)

☐ On Site Testing

6. Test Standard/Method : KOUVA AS 01 : 2016

7. Test Results : Refer to attached results

Note : 1. This report is limited to samples submitted by the applicant and is prohibited from being used for legal or other reasons of dispute.

2. This document is valid only in its original document, and any reproduced copies and electronic copies are not valid. ("Original" means all the reports provided by the KTL including the security procedures.)

3. You can check the contents of the report by scanning the 2D Barcode below. The identity of original reports can be checked in the "Confirm original report" window of the customer's homepage (customer.ktl.re.kr).

Affirmation	Tested by	Technical Manager
	Name : Park seo-hyun (Signature)	Name : JEON, Yong-Woo (Signature)

2020. 11. 10.

Korea Testing Laboratory

87, Digital-ro 26-gil, Guro-gu, Seoul, KOREA Tel.+82-2-860-1139 Fax. +82-2-860-1689

FP104-07-00



※ 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.

Test Results

1. Test Results

Unit	Reference and Results	Reference value	Airborne Bacteria reduction rate
PLASMA AIR DISINFECTOR		over 80 %	98.8 %

2. Test sample

- (1) Manufacturer : Health Well Medical Inc.
(2) Model name : HPA-36C

3. Test Method and Conditions

Test method	Conditions	Bacteria	Temp.	Humidity	Test chamber	Test time	Operation mode	Remarks
KOUVA AS 01 : 2016, Air sterilizer		<i>S.epidermi</i> <i>dis</i> (ATCC 12228)	(23 ± 2) ℃	(50 ± 5) % R.H.	60 m ³	60 min	Plasma	Remove filter

4. Unit Photograph



- End of document -

